

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 003438  
10:15 13/04/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:53  
PRESSURIZING 1:59  
STERILIZING 2:56  
136.0 C 220kPa 2:56  
136.4 C 224kPa 3:26  
136.4 C 222kPa 3:56  
135.9 C 219kPa 4:26  
136.4 C 223kPa 4:56  
136.4 C 223kPa 5:26  
136.1 C 217kPa 5:56  
136.3 C 222kPa 6:26  
136.3 C 222kPa 6:26  
VENTING 6:26  
AIR DRYING 7:18  
CYCLE COMPLETE 23:20

-----  
Digital Signature #  
5AEB99D1F88558D2  
-----