

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2922

Name: RITTER2  
Sterilizer ID: V2313228  
09/25/2023 03:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.4 F 0.0  
2:00 178.9 F 3.2  
4:00 191.6 F 7.4  
6:00 207.7 F 12.2  
8:00 222.2 F 7.0  
10:00 233.0 F 12.0  
12:00 235.3 F 9.7  
14:00 237.6 F 9.5  
16:00 248.6 F 15.0  
18:00 259.9 F 21.4  
20:00 270.5 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 272.2 F 29.5  
2:00 270.9 F 28.9  
3:00 272.6 F 30.3  
4:00 272.1 F 29.6

Min 270.9 F 28.8  
Max 273.3 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.6 F 0.0  
5:00 216.8 F 0.0  
10:00 221.0 F 0.0  
15:00 217.1 F 0.0  
20:00 212.3 F 0.0  
25:00 210.1 F 0.0  
30:00 208.1 F 0.0  
Drying Complete

FILLING: 1:55  
HEATING: 20:49  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:00:24  
CYCLE COMPLETE

09/25/2023 04:55 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_