

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1770

Name: RITTER2  
Sterilizer ID: V2313228  
11/23/2022 02:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 206.6 F 0.0  
2:00 200.2 F 4.4  
4:00 212.6 F 10.5  
6:00 224.8 F 7.0  
8:00 239.5 F 13.5  
10:00 241.9 F 12.3  
12:00 242.8 F 12.1  
14:00 257.0 F 19.6  
16:00 269.3 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.5 F 29.6  
2:00 271.6 F 28.8  
3:00 272.2 F 29.8  
4:00 272.4 F 29.4  
  
Min 271.4 F 28.8  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.1 F 0.0  
5:00 222.4 F 0.0  
10:00 231.9 F 0.0  
15:00 233.0 F 0.0  
20:00 228.0 F 0.0  
25:00 226.6 F 0.0  
30:00 226.6 F 0.0  
Drying Complete

FILLING: 2:25  
HEATING: 17:01  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:56:36  
CYCLE COMPLETE

11/23/2022 02:56 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_