

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1781

Name: RITTER 1  
Sterilizer ID: V2178198  
04/01/2022 10:07 AM

BEGIN UNWRAPPED CYCLE  
Temp: 270 Degrees F  
Time: 3 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 107.5 F 0.0  
2:00 109.0 F 0.9  
4:00 129.2 F 3.6  
6:00 163.7 F 7.6  
8:00 189.7 F 12.5  
10:00 213.2 F 4.8  
12:00 226.2 F 10.2  
14:00 229.9 F 7.2  
16:00 242.3 F 12.5  
18:00 240.4 F 10.6  
20:00 253.7 F 17.4  
22:00 265.5 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 272.3 F 28.7  
2:00 272.9 F 29.3  
3:00 271.7 F 28.6

Min 271.5 F 28.3  
Max 273.0 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.0 F 0.0  
5:00 222.6 F 0.0  
10:00 216.1 F 0.0  
Dry Cycle Aborted

FILLING: 1:13  
HEATING: 23:41  
STERILIZING: 3:00  
VENTING: 2:50  
DRYING: 10:56  
TOTAL CYCLE: 00:41:40

04/01/2022 10:48 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_