
Midmark M11 - v1.0.0
Total Cycles: 1201

Name: RITTER 1
Sterilizer ID: V2178198
09/23/2021 12:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.1 F 0.0
2:00 164.2 F 2.6
4:00 190.4 F 6.7
6:00 215.2 F 11.7
8:00 223.8 F 6.3
10:00 237.1 F 11.5
12:00 233.9 F 8.5
14:00 240.0 F 10.4
16:00 246.3 F 13.4
18:00 259.2 F 20.3
20:00 270.5 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.8 F 29.2
2:00 272.1 F 28.5
3:00 271.5 F 28.5
4:00 272.7 F 29.3

Min 271.2 F 28.2
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 228.2 F 0.0
10:00 215.3 F 0.0
15:00 210.0 F 0.0
20:00 207.3 F 0.0
25:00 212.0 F 0.0
30:00 212.2 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 20:53
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:00:04
CYCLE COMPLETE

09/23/2021 01:18 PM
ACCEPTED: X
REJECTED:
Operator: U02