
Midmark M11 - v1.0.0
Total Cycles: 2056

Name: RITTER 1
Sterilizer ID: V2178198
06/24/2022 03:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.3 F 0.0
2:00 171.1 F 3.2
4:00 194.6 F 7.4
6:00 215.8 F 12.3
8:00 227.6 F 6.8
10:00 234.1 F 12.0
12:00 235.9 F 9.0
14:00 238.8 F 9.9
16:00 246.9 F 13.6
18:00 259.0 F 20.1
20:00 269.7 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.4 F 29.0
2:00 272.6 F 28.9
3:00 272.0 F 28.4
4:00 271.2 F 28.2

Min 271.1 F 28.1
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.7 F 0.0
5:00 230.7 F 0.0
10:00 228.3 F 0.0
15:00 228.6 F 0.0
20:00 228.7 F 0.0
25:00 223.3 F 0.0
30:00 206.1 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 21:06
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:37
CYCLE COMPLETE

06/24/2022 04:58 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____