
Midmark M11 - v1.0.0
Total Cycles: 3854

Name: RITTER 1
Sterilizer ID: V2178198
10/03/2023 10:25 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.3 F 0.0
2:00 163.2 F 1.9
4:00 178.8 F 6.4
6:00 203.9 F 12.5
8:00 224.0 F 7.0
10:00 237.9 F 11.4
12:00 243.0 F 12.3
14:00 242.6 F 11.3
16:00 257.4 F 19.0
18:00 259.6 F 17.4
20:00 270.4 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 27.9
1:00 271.8 F 27.8
2:00 272.1 F 27.9
3:00 272.4 F 28.2

Min 271.7 F 27.8
Max 272.9 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 225.1 F 0.0
5:00 208.9 F 0.0
Dry Cycle Aborted

FILLING: 0:58
HEATING: 20:40
STERILIZING: 3:00
VENTING: 3:02
DRYING: 9:24
TOTAL CYCLE: 00:37:04

10/03/2023 11:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____