
Midmark M11 - v1.0.5
Total Cycles: 513

Name: RITTER2
Sterilizer ID: V2313228
11/02/2021 10:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.6 F 0.0
2:00 182.1 F 2.6
4:00 184.1 F 6.9
6:00 206.5 F 11.9
8:00 222.7 F 7.0
10:00 234.4 F 12.0
12:00 235.4 F 9.6
14:00 237.6 F 9.5
16:00 248.1 F 14.8
18:00 259.9 F 21.3
20:00 270.4 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.7 F 29.5
2:00 271.4 F 29.1
3:00 272.4 F 29.6
4:00 271.4 F 28.9

Min 271.3 F 28.9
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 233.6 F 0.0
10:00 227.8 F 0.0
15:00 225.9 F 0.0
20:00 224.5 F 0.0
25:00 230.5 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 20:48
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:50
CYCLE COMPLETE

11/02/2021 11:22 AM
ACCEPTED: X
REJECTED:
Operator: U01