
Midmark M11 - v1.0.0
Total Cycles: 4854

Name: RITTER 1
Sterilizer ID: V2178198
08/27/2024 10:44 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.0 F 0.0
2:00 170.2 F 0.9
4:00 189.6 F 6.4
6:00 211.9 F 12.8
8:00 225.8 F 7.8
10:00 235.6 F 9.2
12:00 244.7 F 12.9
14:00 243.6 F 11.8
16:00 258.2 F 19.4
18:00 270.2 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 271.7 F 27.9
2:00 271.9 F 28.0
3:00 272.2 F 28.2

Min 271.6 F 27.9
Max 272.9 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.4 F 0.0
Dry Cycle Aborted

FILLING: 2:08
HEATING: 18:44
STERILIZING: 3:00
VENTING: 2:60
DRYING: 4:09
TOTAL CYCLE: 00:31:01

08/27/2024 11:13 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____