
Midmark M11 - v1.0.5
Total Cycles: 809

Name: RITTER2
Sterilizer ID: V2313228
02/16/2022 01:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.8 F 0.0
2:00 140.3 F 2.3
4:00 165.1 F 6.0
6:00 190.5 F 11.0
8:00 213.0 F 4.5
10:00 224.2 F 10.1
12:00 232.4 F 7.9
14:00 242.4 F 13.2
16:00 242.8 F 12.2
18:00 255.9 F 19.1
20:00 267.5 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.6 F 29.9
2:00 271.9 F 29.3
3:00 271.3 F 29.3
4:00 272.5 F 30.0

Min 271.2 F 28.9
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.1 F 0.0
5:00 226.9 F 0.0
10:00 211.5 F 0.0
15:00 217.1 F 0.0
20:00 220.5 F 0.0
25:00 221.6 F 0.0
30:00 216.8 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 21:24
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:52
CYCLE COMPLETE

02/16/2022 02:01 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____