
Midmark M11 - v1.0.5
Total Cycles: 1357

Name: RITTER2
Sterilizer ID: V2313228
07/27/2022 01:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.0 F 0.0
2:00 168.3 F 0.1
4:00 183.1 F 1.9
6:00 210.3 F 6.6
8:00 229.8 F 12.0
10:00 230.9 F 8.4
12:00 241.3 F 11.5
14:00 242.9 F 12.7
16:00 242.8 F 12.0
18:00 255.9 F 18.9
20:00 267.4 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.4 F 29.4
2:00 271.5 F 29.0
3:00 272.4 F 29.4
4:00 271.6 F 28.8

Min 271.4 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 229.0 F 0.0
10:00 234.7 F 0.0
15:00 230.1 F 0.0
20:00 226.4 F 0.0
25:00 232.6 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 21:24
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:36
CYCLE COMPLETE

07/27/2022 02:30 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____