
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007464
8:18 30/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:20
PRESSURIZING 3:22
STERILIZING 4:25
136.0 C 220kPa 4:25
136.4 C 224kPa 4:55
136.4 C 222kPa 5:25
136.4 C 221kPa 5:55
136.2 C 222kPa 6:25
136.4 C 222kPa 6:55
136.4 C 224kPa 7:25
136.4 C 223kPa 7:55
136.4 C 223kPa 7:55
VENTING 7:55
AIR DRYING 8:46
CYCLE COMPLETE 9:05

Digital Signature #
0EF62A42A18BC890
