
Midmark M11 - v1.0.0
Total Cycles: 2055

Name: RITTER 1
Sterilizer ID: V2178198
06/24/2022 02:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.1 F 0.0
2:00 189.9 F 3.7
4:00 201.5 F 7.9
6:00 220.3 F 13.4
8:00 227.2 F 7.5
10:00 239.4 F 12.6
12:00 237.2 F 9.8
14:00 238.1 F 8.5
16:00 248.8 F 14.6
18:00 260.6 F 21.2
20:00 271.5 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.7 F 29.0
2:00 271.6 F 28.2
3:00 272.4 F 29.0
4:00 272.2 F 28.7

Min 271.2 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.0 F 0.0
5:00 227.5 F 0.0
10:00 241.2 F 0.0
15:00 227.0 F 0.0
20:00 228.9 F 0.0
25:00 228.9 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 20:40
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:14
CYCLE COMPLETE

06/24/2022 03:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____