
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.3uS / 0.8ppm
CYCLE NUMBER 000621
12:49 01/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:29
PRESSURIZING 3:31
STERILIZING 4:46
136.0 C 221kPa 4:46
135.9 C 228kPa 5:17
136.4 C 232kPa 5:47
136.1 C 229kPa 6:17
136.0 C 229kPa 6:47
136.4 C 232kPa 7:17
136.4 C 231kPa 7:47
136.4 C 232kPa 8:17
136.4 C 232kPa 8:17
VENTING 8:17
AIR DRYING 9:09
CYCLE COMPLETE 24:09

Digital Signature #
5252E44A784B53AE
