

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1428

Name: RITTER2  
Sterilizer ID: V2313228  
08/29/2022 02:48 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 206.5 F 0.0  
2:00 201.7 F 3.2  
4:00 206.4 F 8.9  
6:00 221.7 F 8.5  
8:00 233.9 F 11.3  
10:00 235.6 F 9.4  
12:00 237.6 F 9.3  
14:00 249.8 F 15.7  
16:00 261.9 F 22.7  
18:00 272.8 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.1  
1:00 272.4 F 29.6  
2:00 271.5 F 29.3  
3:00 272.6 F 29.7  
4:00 271.5 F 29.0  
  
Min 271.4 F 29.0  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.4 F 0.0  
5:00 218.0 F 0.0  
10:00 236.1 F 0.0  
15:00 239.0 F 0.0  
20:00 231.3 F 0.0  
25:00 223.2 F 0.0  
30:00 220.6 F 0.0  
  
Drying Complete

FILLING: 2:18  
HEATING: 18:21  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:57:50  
CYCLE COMPLETE

08/29/2022 03:45 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_