
Midmark M11 - v1.0.5
Total Cycles: 1874

Name: RITTER2
Sterilizer ID: V2313228
12/21/2022 02:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 218.9 F 0.0
2:00 209.2 F 4.4
4:00 211.4 F 10.3
6:00 222.1 F 6.1
8:00 235.8 F 12.6
10:00 239.2 F 11.0
12:00 239.2 F 10.4
14:00 253.0 F 17.3
16:00 264.9 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.6 F 29.5
2:00 271.6 F 28.7
3:00 272.3 F 29.7
4:00 272.1 F 29.2

Min 271.4 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.3 F 0.0
5:00 227.2 F 0.0
10:00 234.1 F 0.0
15:00 229.0 F 0.0
20:00 227.4 F 0.0
25:00 228.0 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 2:35
HEATING: 17:49
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:35
CYCLE COMPLETE

12/21/2022 03:54 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____