

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1808

Name: RITTER2  
Sterilizer ID: V2313228  
12/06/2022 07:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 81.9 F 0.0  
2:00 82.8 F 0.0  
4:00 164.7 F 0.0  
6:00 169.2 F 0.0  
8:00 212.8 F 1.0  
10:00 216.8 F 5.0  
12:00 220.8 F 9.0  
14:00 233.3 F 13.3  
16:00 233.6 F 8.5  
18:00 244.0 F 13.2  
20:00 239.7 F 10.5  
22:00 235.5 F 8.0  
24:00 249.2 F 14.8  
26:00 260.3 F 21.2  
28:00 270.5 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 29.7  
1:00 271.9 F 29.9  
2:00 272.0 F 30.0  
3:00 272.2 F 30.0  
4:00 272.1 F 30.0

Min 271.0 F 29.6  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.2 F 0.0  
5:00 262.3 F 0.0  
10:00 231.3 F 0.0  
15:00 229.3 F 0.0  
20:00 222.1 F 0.0  
25:00 207.2 F 0.0  
30:00 209.6 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 29:31  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 01:07:25  
CYCLE COMPLETE

12/06/2022 08:52 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_