
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 002675
11:50 11/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:35
PRESSURIZING 2:37
STERILIZING 3:34
136.4 C 221kPa 3:34
136.2 C 223kPa 4:04
136.4 C 223kPa 4:34
136.5 C 223kPa 5:04
135.9 C 218kPa 5:34
136.4 C 223kPa 6:04
136.4 C 225kPa 6:34
136.2 C 223kPa 7:04
136.2 C 223kPa 7:04
VENTING 7:04
AIR DRYING 7:58
CYCLE COMPLETE 7:59

Digital Signature #
7BB4156861400AED
