
Midmark M11 - v1.0.5
Total Cycles: 3275

Name: RITTER2
Sterilizer ID: V2313228
01/04/2024 10:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.7 F 0.0
2:00 152.3 F 0.9
4:00 160.6 F 3.4
6:00 180.7 F 7.1
8:00 201.4 F 11.5
10:00 223.2 F 5.5
12:00 226.3 F 10.1
14:00 232.4 F 8.4
16:00 241.1 F 12.5
18:00 241.3 F 11.2
20:00 253.6 F 17.5
22:00 264.6 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.9
1:00 272.7 F 29.9
2:00 271.7 F 29.5
3:00 272.3 F 29.5
4:00 272.5 F 30.0

Min 271.0 F 28.7
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.9 F 0.0
5:00 242.6 F 0.0
10:00 194.5 F 0.0
15:00 189.7 F 0.0
20:00 185.0 F 0.0
25:00 182.5 F 0.0
30:00 180.6 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 23:59
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:02:35
CYCLE COMPLETE

01/04/2024 11:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____