

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1719

Name: RITTER 1  
Sterilizer ID: V2178198  
03/15/2022 12:42 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 151.0 F 0.0  
2:00 152.8 F 1.2  
4:00 170.8 F 4.0  
6:00 190.8 F 7.8  
8:00 215.6 F 12.8  
10:00 221.7 F 6.5  
12:00 233.8 F 11.1  
14:00 232.5 F 7.5  
16:00 244.8 F 13.6  
18:00 243.5 F 12.0  
20:00 256.2 F 18.6  
22:00 267.4 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.5 F 29.1  
2:00 272.0 F 28.4  
3:00 271.5 F 28.4  
4:00 272.5 F 29.1

Min 271.3 F 28.1  
Max 272.5 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.7 F 0.0  
5:00 198.4 F 0.0  
10:00 198.2 F 0.0  
15:00 197.0 F 0.0  
20:00 215.8 F 0.0  
25:00 221.6 F 0.0  
30:00 223.4 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 23:27  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:01:36  
CYCLE COMPLETE

03/15/2022 01:43 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_