

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4625

Name: RITTER 1  
Sterilizer ID: V2178198  
06/11/2024 09:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 125.2 F 0.0  
2:00 137.3 F 1.2  
4:00 145.6 F 3.9  
6:00 173.2 F 6.8  
8:00 201.7 F 10.6  
10:00 210.4 F 4.5  
12:00 223.7 F 6.2  
14:00 233.5 F 10.1  
16:00 235.0 F 9.0  
18:00 239.0 F 10.7  
20:00 236.7 F 8.1  
22:00 248.0 F 14.2  
24:00 259.0 F 20.2  
26:00 269.1 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.7 F 29.0  
2:00 272.0 F 28.3  
3:00 271.9 F 28.7  
4:00 272.7 F 28.9

Min 271.5 F 28.2  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.3 F 0.0  
5:00 219.7 F 0.0  
10:00 223.4 F 0.0  
15:00 231.7 F 0.0  
20:00 228.7 F 0.0  
25:00 227.1 F 0.0  
30:00 225.8 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 27:10  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:05:44  
CYCLE COMPLETE

06/11/2024 11:00 AM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_