
Midmark M11 - v1.0.5
Total Cycles: 3943

Name: RITTER2
Sterilizer ID: V2313228
09/13/2024 10:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.6 F 0.0
2:00 163.2 F 1.4
4:00 164.9 F 4.3
6:00 187.2 F 8.1
8:00 207.6 F 12.7
10:00 217.7 F 6.2
12:00 229.5 F 10.5
14:00 231.7 F 8.0
16:00 241.9 F 12.6
18:00 240.8 F 11.2
20:00 252.7 F 17.1
22:00 263.1 F 23.3
24:00 272.7 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.4 F 29.5
2:00 271.4 F 28.9
3:00 272.4 F 29.7
4:00 271.9 F 29.0

Min 271.3 F 28.8
Max 272.4 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 209.6 F 0.0
10:00 206.9 F 0.0
15:00 200.5 F 0.0
20:00 200.3 F 0.0
25:00 200.6 F 0.0
30:00 202.9 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 24:23
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:03:11
CYCLE COMPLETE

09/13/2024 11:25 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____