

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2184

Name: RITTER2  
Sterilizer ID: V2313228  
03/28/2023 09:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 154.0 F 0.0  
2:00 153.9 F 2.3  
4:00 157.8 F 5.2  
6:00 189.8 F 9.8  
8:00 209.1 F 6.2  
10:00 223.6 F 9.4  
12:00 232.2 F 9.5  
14:00 238.7 F 11.8  
16:00 238.1 F 10.1  
18:00 251.4 F 16.7  
20:00 263.1 F 23.5  
22:00 272.0 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.2  
1:00 272.4 F 29.9  
2:00 271.6 F 29.2  
3:00 272.0 F 29.9  
4:00 272.1 F 29.5

Min 271.1 F 29.0  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.2 F 0.0  
5:00 204.8 F 0.0  
10:00 216.1 F 0.0  
15:00 221.9 F 0.0  
20:00 224.3 F 0.0  
25:00 224.5 F 0.0  
30:00 223.7 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 22:02  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:00:15  
CYCLE COMPLETE

03/28/2023 10:14 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_