
Midmark M11 - v1.0.5
Total Cycles: 2748

Name: RITTER2
Sterilizer ID: V2313228
08/10/2023 07:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 71.5 F 0.0
4:00 83.6 F 0.0
6:00 104.0 F 0.0
8:00 154.8 F 0.0
10:00 176.7 F 0.0
12:00 200.9 F 0.0
14:00 211.6 F 2.7
16:00 219.8 F 5.5
18:00 228.1 F 8.6
20:00 237.0 F 12.3
22:00 235.0 F 9.1
24:00 241.9 F 11.5
26:00 244.6 F 12.9
28:00 242.9 F 12.0
30:00 254.9 F 18.1
32:00 265.3 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.6 F 29.2
2:00 271.7 F 28.9
3:00 272.5 F 29.1
4:00 271.8 F 28.9

Min 271.5 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.4 F 0.0
5:00 213.0 F 0.0
10:00 207.6 F 0.0
15:00 205.2 F 0.0
20:00 203.9 F 0.0
25:00 204.2 F 0.0
30:00 207.5 F 0.0

Drying Complete

FILLING: 0:48
HEATING: 33:54
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:12:03
CYCLE COMPLETE

08/10/2023 09:07 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____