

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1237

Name: RITTER2  
Sterilizer ID: V2313228  
06/28/2022 03:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 112.6 F 0.0  
2:00 115.8 F 1.0  
4:00 136.7 F 3.7  
6:00 166.1 F 7.3  
8:00 185.0 F 11.2  
10:00 210.6 F 3.8  
12:00 221.1 F 7.7  
14:00 232.3 F 12.0  
16:00 233.0 F 8.5  
18:00 243.6 F 12.4  
20:00 243.4 F 12.3  
22:00 255.0 F 18.3  
24:00 265.4 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.6 F 29.5  
2:00 271.7 F 28.8  
3:00 272.0 F 29.5  
4:00 272.8 F 29.4

Min 271.3 F 28.7  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.3 F 0.0  
5:00 230.4 F 0.0  
10:00 208.5 F 0.0  
15:00 218.3 F 0.0  
20:00 197.4 F 0.0  
25:00 197.2 F 0.0  
30:00 196.3 F 0.0  
Drying Complete

FILLING: 1:20  
HEATING: 25:55  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:04:28  
CYCLE COMPLETE

06/28/2022 04:04 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_