

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2530

Name: RITTER 1  
Sterilizer ID: V2178198  
11/08/2022 01:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.8 F 0.0  
2:00 177.6 F 2.5  
4:00 200.2 F 6.3  
6:00 212.7 F 11.0  
8:00 210.5 F 4.4  
10:00 228.0 F 9.6  
12:00 232.6 F 8.5  
14:00 240.6 F 11.9  
16:00 238.7 F 9.9  
18:00 252.5 F 16.6  
20:00 264.3 F 23.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.3  
1:00 272.4 F 29.0  
2:00 271.3 F 28.2  
3:00 272.0 F 29.1  
4:00 272.3 F 28.7

Min 271.1 F 28.1  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.0 F 0.0  
5:00 235.0 F 0.0  
10:00 238.3 F 0.0  
15:00 235.1 F 0.0  
20:00 224.6 F 0.0  
25:00 227.4 F 0.0  
30:00 225.3 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 21:58  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:59:59  
CYCLE COMPLETE

11/08/2022 02:57 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_