
Midmark M11 - v1.0.0
Total Cycles: 909

Name: RITTER 1
Sterilizer ID: V2178198
06/17/2021 08:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.6 F 0.0
2:00 143.1 F 0.9
4:00 148.4 F 2.7
6:00 170.6 F 5.8
8:00 211.6 F 10.0
10:00 226.5 F 5.3
12:00 222.0 F 8.5
14:00 233.9 F 13.2
16:00 236.1 F 10.1
18:00 236.7 F 8.6
20:00 247.9 F 14.3
22:00 259.5 F 20.6
24:00 270.0 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.6 F 29.2
2:00 271.8 F 28.6
3:00 271.3 F 28.6
4:00 272.7 F 29.2

Min 271.1 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.7 F 0.0
5:00 214.8 F 0.0
10:00 213.8 F 0.0
15:00 212.2 F 0.0
20:00 236.2 F 0.0
25:00 231.3 F 0.0
30:00 230.2 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 24:59
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:03:21
CYCLE COMPLETE

06/17/2021 09:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____