
Midmark M11 - v1.0.0
Total Cycles: 3148

Name: RITTER 1
Sterilizer ID: V2178198
04/21/2023 05:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.7 F 0.0
2:00 165.1 F 0.0
4:00 198.3 F 0.8
6:00 215.6 F 5.8
8:00 229.0 F 10.8
10:00 231.9 F 6.9
12:00 243.2 F 13.0
14:00 242.0 F 11.5
16:00 240.7 F 10.6
18:00 254.9 F 17.7
20:00 266.9 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.3 F 28.4
2:00 272.2 F 28.7
3:00 272.0 F 28.2
4:00 272.4 F 28.8

Min 271.5 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 204.0 F 0.0
10:00 222.7 F 0.0
15:00 235.8 F 0.0
20:00 227.9 F 0.0
25:00 225.7 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 21:23
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:39
CYCLE COMPLETE

04/21/2023 06:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____