
Midmark M11 - v1.0.0
Total Cycles: 3086

Name: RITTER 1
Sterilizer ID: V2178198
04/10/2023 05:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.8 F 0.0
2:00 174.1 F 3.2
4:00 188.8 F 7.2
6:00 211.0 F 12.2
8:00 223.4 F 7.2
10:00 236.1 F 12.6
12:00 236.7 F 9.9
14:00 236.4 F 8.5
16:00 249.6 F 15.1
18:00 261.8 F 22.0
20:00 272.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.4 F 28.9
2:00 271.6 F 28.6
3:00 272.4 F 29.0
4:00 271.4 F 28.4

Min 271.4 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.1 F 0.0
5:00 236.2 F 0.0
10:00 228.8 F 0.0
15:00 231.0 F 0.0
20:00 225.5 F 0.0
25:00 227.6 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 20:22
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:17
CYCLE COMPLETE

04/10/2023 06:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____