
Midmark M11 - v1.0.0
Total Cycles: 2554

Name: RITTER 1
Sterilizer ID: V2178198
11/14/2022 03:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.6 F 0.0
2:00 185.4 F 4.5
4:00 198.2 F 9.3
6:00 220.6 F 6.7
8:00 232.6 F 10.0
10:00 235.8 F 7.9
12:00 244.8 F 13.0
14:00 245.0 F 12.7
16:00 258.2 F 19.9
18:00 270.1 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.5 F 28.9
2:00 271.9 F 28.7
3:00 272.4 F 28.9
4:00 271.7 F 28.5

Min 271.6 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.2 F 0.0
5:00 222.0 F 0.0
10:00 231.7 F 0.0
15:00 230.5 F 0.0
20:00 229.9 F 0.0
25:00 228.5 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 18:52
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:51
CYCLE COMPLETE

11/14/2022 04:45 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____