
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001046
12:20 28/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:45
PRESSURIZING 4:10
STERILIZING 6:08
136.4 C 224kPa 6:08
136.1 C 223kPa 6:38
136.2 C 224kPa 7:08
136.4 C 225kPa 7:38
136.2 C 222kPa 8:08
136.4 C 226kPa 8:38
136.2 C 223kPa 9:08
136.4 C 225kPa 9:38
136.4 C 225kPa 9:38
VENTING 9:38
AIR DRYING 10:34
CYCLE COMPLETE 20:05

Digital Signature #
4C9D9265D2D70412
