
Midmark M11 - v1.0.0
Total Cycles: 4229

Name: RITTER 1
Sterilizer ID: V2178198
01/24/2024 07:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.4 F 0.0
2:00 181.6 F 4.6
4:00 205.4 F 10.5
6:00 223.9 F 6.2
8:00 238.5 F 12.5
10:00 239.5 F 10.4
12:00 237.8 F 9.2
14:00 253.7 F 16.9
16:00 266.7 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.0
1:00 271.6 F 27.8
2:00 272.1 F 28.2
3:00 272.6 F 28.6
4:00 272.2 F 28.5

Min 271.6 F 27.8
Max 272.8 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.6 F 0.0
5:00 224.1 F 0.0
10:00 227.5 F 0.0
15:00 246.0 F 0.0
20:00 229.7 F 0.0
25:00 226.8 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 2:25
HEATING: 17:20
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:56:54
CYCLE COMPLETE

01/24/2024 08:07 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____