
Midmark M11 - v1.0.5
Total Cycles: 404

Name: RITTER2
Sterilizer ID: V2313228
09/29/2021 06:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.8 F 0.0
2:00 74.8 F 0.0
4:00 75.3 F 0.0
6:00 85.9 F 0.1
8:00 132.7 F 1.2
10:00 155.8 F 4.4
12:00 175.5 F 7.7
14:00 192.9 F 11.1
16:00 217.6 F 6.5
18:00 235.0 F 9.9
20:00 230.1 F 10.2
22:00 241.8 F 11.5
24:00 238.1 F 9.7
26:00 250.4 F 15.6
28:00 260.7 F 21.5
30:00 270.1 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.3
1:00 271.8 F 29.8
2:00 272.4 F 30.1
3:00 272.5 F 29.8
4:00 272.1 F 29.6

Min 270.9 F 29.3
Max 273.3 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.9 F 0.0
5:00 233.3 F 0.0
10:00 225.7 F 0.0
15:00 224.4 F 0.0
20:00 221.2 F 0.0
25:00 221.5 F 0.0
30:00 218.2 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 31:34
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:09:42
CYCLE COMPLETE

09/29/2021 07:32 AM
ACCEPTED: X
REJECTED:
Operator: U01