
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001857
7:36 16/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:40
PRESSURIZING 3:42
STERILIZING 5:11
136.4 C 222kPa 5:11
136.1 C 222kPa 5:41
136.5 C 222kPa 6:11
136.1 C 222kPa 6:41
136.4 C 225kPa 7:11
136.4 C 221kPa 7:41
136.2 C 223kPa 8:11
136.2 C 224kPa 8:41
136.2 C 224kPa 8:41
VENTING 8:41
AIR DRYING 9:32
CYCLE COMPLETE 23:46

Digital Signature #
39DC2EA9C108A838
