

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001545  
12:50 24/10/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:35  
PRESSURIZING 2:39  
STERILIZING 3:38  
136.5 C 224kPa 3:38  
136.4 C 227kPa 4:08  
136.4 C 224kPa 4:38  
136.2 C 223kPa 5:08  
136.3 C 222kPa 5:38  
136.4 C 226kPa 6:08  
136.1 C 223kPa 6:38  
136.3 C 226kPa 7:08  
136.3 C 226kPa 7:08  
VENTING 7:08  
AIR DRYING 8:03  
CYCLE COMPLETE 9:47  
-----

Digital Signature #  
7B12A810AAF4D45F  
-----