
Midmark M11 - v1.0.0
Total Cycles: 759

Name: RITTER 1
Sterilizer ID: V2178198
03/31/2021 09:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.8 F 0.0
2:00 168.7 F 2.1
4:00 174.1 F 6.2
6:00 196.2 F 11.2
8:00 223.8 F 6.0
10:00 226.8 F 11.7
12:00 236.3 F 8.8
14:00 238.6 F 9.5
16:00 247.5 F 13.7
18:00 260.3 F 20.8
20:00 271.8 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 272.1 F 29.1
2:00 272.5 F 28.9
3:00 272.3 F 28.8
4:00 272.2 F 28.7

Min 271.5 F 28.5
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.1 F 0.0
5:00 233.0 F 0.0
10:00 226.5 F 0.0
15:00 206.8 F 0.0
20:00 208.6 F 0.0
25:00 220.4 F 0.0
Dry Cycle Aborted

FILLING: 1:06
HEATING: 21:05
STERILIZING: 4:00
VENTING: 2:57
DRYING: 29:03
TOTAL CYCLE: 00:58:10

03/31/2021 10:18 AM
ACCEPTED: X
REJECTED:
Operator: U01