
Midmark M11 - v1.0.0
Total Cycles: 967

Name: RITTER 1
Sterilizer ID: V2178198
07/06/2021 10:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 215.4 F 0.0
2:00 201.5 F 3.7
4:00 207.3 F 8.7
6:00 221.8 F 8.7
8:00 232.6 F 9.1
10:00 238.3 F 9.9
12:00 243.7 F 12.3
14:00 241.8 F 11.0
16:00 254.9 F 17.7
18:00 266.2 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.1
1:00 272.6 F 28.7
2:00 271.1 F 28.1
3:00 272.4 F 28.7
4:00 271.3 F 28.0

Min 271.0 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.8 F 0.0
5:00 227.9 F 0.0
10:00 205.7 F 0.0
15:00 202.5 F 0.0
20:00 209.0 F 0.0
25:00 218.1 F 0.0
30:00 218.5 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 19:39
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:40
CYCLE COMPLETE

07/06/2021 11:17 AM
ACCEPTED: X
REJECTED:
Operator: U01