

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 005136  
12:46 16/02/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:04  
PRESSURIZING 2:08  
STERILIZING 3:10  
136.4 C 220kPa 3:10  
136.3 C 223kPa 3:40  
136.2 C 220kPa 4:10  
136.4 C 223kPa 4:40  
136.3 C 224kPa 5:10  
136.3 C 223kPa 5:40  
136.4 C 217kPa 6:10  
136.4 C 222kPa 6:40  
136.4 C 222kPa 6:40  
VENTING 6:40  
AIR DRYING 7:40  
CYCLE COMPLETE 12:03  
-----

Digital Signature #  
F6798EEFE986F998  
-----