
Midmark M11 - v1.0.5
Total Cycles: 2324

Name: RITTER2
Sterilizer ID: V2313228
05/02/2023 10:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 211.0 F 0.0
2:00 201.8 F 3.4
4:00 204.7 F 9.0
6:00 222.3 F 8.5
8:00 232.9 F 10.9
10:00 233.5 F 8.9
12:00 237.8 F 9.8
14:00 248.7 F 15.3
16:00 261.2 F 22.4
18:00 272.7 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.2 F 29.5
2:00 272.0 F 29.9
3:00 271.9 F 29.3
4:00 272.3 F 30.0

Min 271.4 F 29.1
Max 272.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.5 F 0.0
5:00 211.6 F 0.0
10:00 221.2 F 0.0
15:00 222.6 F 0.0
20:00 222.4 F 0.0
25:00 223.4 F 0.0
30:00 222.7 F 0.0
Drying Complete

FILLING: 2:16
HEATING: 18:22
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:57:56
CYCLE COMPLETE

05/02/2023 10:59 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____