
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008413
10:10 13/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:25
PRESSURIZING 2:28
STERILIZING 3:37
136.5 C 223kPa 3:37
136.1 C 219kPa 4:07
136.2 C 222kPa 4:37
136.4 C 224kPa 5:07
136.4 C 220kPa 5:37
136.4 C 224kPa 6:07
136.4 C 221kPa 6:37
136.1 C 221kPa 7:07
136.1 C 221kPa 7:07
VENTING 7:08
AIR DRYING 7:58
CYCLE COMPLETE 37:58

Digital Signature #
057A0A62F7D09976
