
Midmark M11 - v1.0.0
Total Cycles: 769

Name: RITTER 1
Sterilizer ID: V2178198
04/06/2021 08:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.6 F 0.0
2:00 188.5 F 2.5
4:00 183.0 F 6.8
6:00 219.1 F 12.5
8:00 221.5 F 7.6
10:00 241.3 F 13.6
12:00 240.6 F 11.4
14:00 239.0 F 9.8
16:00 253.6 F 17.0
18:00 265.7 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.5 F 29.2
2:00 272.1 F 29.2
3:00 272.2 F 29.2
4:00 272.0 F 29.2

Min 270.6 F 28.0
Max 273.1 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 238.7 F 0.0
10:00 228.8 F 0.0
15:00 229.6 F 0.0
20:00 211.0 F 0.0
25:00 197.6 F 0.0
Dry Cycle Aborted

FILLING: 1:12
HEATING: 19:30
STERILIZING: 4:00
VENTING: 2:56
DRYING: 29:01
TOTAL CYCLE: 00:56:39

04/06/2021 09:46 AM
ACCEPTED: X
REJECTED:
Operator: U01