
Midmark M11 - v1.0.0
Total Cycles: 3251

Name: RITTER 1
Sterilizer ID: V2178198
05/17/2023 12:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.1 F 0.0
2:00 188.5 F 2.5
4:00 206.2 F 7.8
6:00 222.5 F 13.1
8:00 231.2 F 8.7
10:00 237.7 F 10.3
12:00 243.9 F 12.5
14:00 242.7 F 11.4
16:00 256.2 F 18.5
18:00 268.0 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.4 F 28.8
2:00 271.6 F 28.1
3:00 272.0 F 28.8
4:00 272.3 F 28.6

Min 271.3 F 28.0
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.5 F 0.0
5:00 229.6 F 0.0
10:00 233.5 F 0.0
15:00 232.2 F 0.0
20:00 225.9 F 0.0
25:00 226.0 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 19:17
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

05/17/2023 01:11 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____