
Midmark M11 - v1.0.5
Total Cycles: 1394

Name: RITTER2
Sterilizer ID: V2313228
08/16/2022 06:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.0 F 0.0
2:00 74.1 F 0.0
4:00 76.3 F 0.0
6:00 148.2 F 0.0
8:00 180.8 F 1.6
10:00 198.9 F 5.0
12:00 210.5 F 8.7
14:00 221.5 F 12.6
16:00 220.4 F 6.3
18:00 229.8 F 10.2
20:00 232.6 F 9.2
22:00 238.7 F 11.1
24:00 235.7 F 8.8
26:00 248.7 F 14.9
28:00 259.4 F 21.0
30:00 269.6 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.0 F 29.4
2:00 272.4 F 29.3
3:00 272.1 F 29.5
4:00 272.4 F 29.3

Min 270.9 F 28.6
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.1 F 0.0
5:00 231.5 F 0.0
10:00 225.3 F 0.0
15:00 192.0 F 0.0
20:00 175.9 F 0.0
25:00 188.8 F 0.0
30:00 216.4 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 30:59
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:09:04
CYCLE COMPLETE

08/16/2022 08:02 AM
ACCEPTED: X
REJECTED:
Operator: U02