
Midmark M11 - v1.0.5
Total Cycles: 2475

Name: RITTER2
Sterilizer ID: V2313228
06/06/2023 03:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.0 F 0.0
2:00 180.7 F 4.0
4:00 191.1 F 8.6
6:00 211.3 F 9.1
8:00 224.9 F 9.3
10:00 233.6 F 9.2
12:00 241.7 F 12.9
14:00 241.9 F 12.0
16:00 255.2 F 18.9
18:00 267.0 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.4
1:00 272.5 F 29.9
2:00 271.6 F 29.5
3:00 272.5 F 29.9
4:00 271.6 F 29.2

Min 271.5 F 29.2
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 223.3 F 0.0
10:00 232.3 F 0.0
15:00 231.7 F 0.0
20:00 230.9 F 0.0
25:00 223.6 F 0.0
30:00 220.4 F 0.0

Drying Complete

FILLING: 1:38
HEATING: 19:26
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:18
CYCLE COMPLETE

06/06/2023 04:47 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____