
Midmark M11 - v1.0.0
Total Cycles: 1907

Name: RITTER 1
Sterilizer ID: V2178198
05/04/2022 10:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.5 F 0.0
2:00 164.7 F 2.1
4:00 174.4 F 5.7
6:00 195.3 F 10.3
8:00 219.9 F 4.9
10:00 223.5 F 9.0
12:00 238.3 F 10.8
14:00 239.0 F 11.1
16:00 237.1 F 9.1
18:00 250.9 F 15.8
20:00 262.6 F 22.6
22:00 272.4 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.6 F 29.2
2:00 271.6 F 28.5
3:00 272.1 F 29.2
4:00 272.6 F 28.9

Min 271.2 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 211.1 F 0.0
10:00 202.7 F 0.0
15:00 207.6 F 0.0
20:00 212.8 F 0.0
25:00 229.6 F 0.0
30:00 235.1 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 22:17
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:21
CYCLE COMPLETE

05/04/2022 11:20 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____