
Midmark M11 - v1.0.0
Total Cycles: 2690

Name: RITTER 1
Sterilizer ID: V2178198
12/19/2022 12:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.4 F 0.0
2:00 199.3 F 3.6
4:00 208.5 F 8.5
6:00 226.5 F 7.5
8:00 233.7 F 10.0
10:00 236.4 F 7.8
12:00 242.6 F 11.7
14:00 245.9 F 13.2
16:00 259.3 F 20.4
18:00 270.7 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.5 F 29.0
2:00 271.4 F 28.7
3:00 272.6 F 29.1
4:00 271.5 F 28.4

Min 271.4 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.6 F 0.0
5:00 233.3 F 0.0
10:00 230.9 F 0.0
15:00 228.9 F 0.0
20:00 227.1 F 0.0
25:00 227.1 F 0.0
30:00 228.6 F 0.0

Drying Complete

FILLING: 1:57
HEATING: 18:45
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:51
CYCLE COMPLETE

12/19/2022 12:57 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____