

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2300

Name: RITTER 1  
Sterilizer ID: V2178198  
09/07/2022 03:54 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 211.5 F 0.0  
2:00 210.5 F 2.8  
4:00 218.5 F 9.5  
6:00 226.7 F 6.8  
8:00 242.3 F 13.0  
10:00 242.4 F 11.5  
12:00 241.2 F 10.7  
14:00 256.3 F 18.3  
16:00 269.1 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.9  
1:00 272.2 F 28.1  
2:00 272.7 F 28.8  
3:00 271.7 F 27.8  
4:00 272.9 F 28.7  
  
Min 271.5 F 27.8  
Max 272.9 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.7 F 0.0  
5:00 216.2 F 0.0  
10:00 232.3 F 0.0  
15:00 228.2 F 0.0  
20:00 222.4 F 0.0  
25:00 218.6 F 0.0  
30:00 216.6 F 0.0  
Drying Complete

FILLING: 2:24  
HEATING: 16:59  
STERILIZING: 4:00  
VENTING: 2:56  
DRYING: 30:00  
TOTAL CYCLE: 00:56:19  
CYCLE COMPLETE

09/07/2022 04:50 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_