
Midmark M11 - v1.0.0
Total Cycles: 3216

Name: RITTER 1
Sterilizer ID: V2178198
05/11/2023 07:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.2 F 0.0
2:00 78.3 F 0.0
4:00 91.0 F 0.0
6:00 127.7 F 0.0
8:00 145.3 F 1.6
10:00 175.2 F 4.6
12:00 196.2 F 8.1
14:00 214.2 F 11.7
16:00 222.0 F 4.7
18:00 230.1 F 8.7
20:00 238.8 F 12.9
22:00 236.4 F 9.4
24:00 239.3 F 9.8
26:00 245.6 F 13.0
28:00 256.9 F 19.0
30:00 267.3 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 271.9 F 28.4
2:00 272.5 F 29.0
3:00 271.3 F 28.2
4:00 272.2 F 28.6

Min 271.2 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.0 F 0.0
5:00 222.7 F 0.0
10:00 228.6 F 0.0
15:00 232.5 F 0.0
20:00 229.0 F 0.0
25:00 231.2 F 0.0
30:00 230.9 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 31:24
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:09:39
CYCLE COMPLETE

05/11/2023 08:18 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____