

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1132

Name: RITTER 1  
Sterilizer ID: V2178198  
08/30/2021 02:35 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 161.2 F 0.0  
2:00 166.5 F 2.9  
4:00 191.6 F 6.8  
6:00 211.0 F 12.2  
8:00 222.3 F 6.9  
10:00 235.8 F 12.1  
12:00 235.4 F 9.2  
14:00 238.3 F 9.3  
16:00 247.7 F 14.4  
18:00 260.2 F 21.2  
20:00 271.5 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.6  
1:00 272.7 F 29.2  
2:00 271.7 F 28.6  
3:00 272.8 F 29.3  
4:00 272.0 F 28.6  
  
Min 271.6 F 28.5  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.5 F 0.0  
5:00 211.1 F 0.0  
10:00 220.4 F 0.0  
15:00 217.8 F 0.0  
20:00 215.8 F 0.0  
25:00 216.3 F 0.0  
30:00 215.4 F 0.0  
Drying Complete

FILLING: 1:31  
HEATING: 20:37  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:59:15  
CYCLE COMPLETE

08/30/2021 03:52 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02