

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3071

Name: RITTER 1  
Sterilizer ID: V2178198  
04/04/2023 08:15 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 79.8 F 0.0  
2:00 81.3 F 0.0  
4:00 100.6 F 0.0  
6:00 126.8 F 0.8  
8:00 156.7 F 2.5  
10:00 181.0 F 4.9  
12:00 197.2 F 8.1  
14:00 210.3 F 11.8  
16:00 218.9 F 4.9  
18:00 231.4 F 9.2  
20:00 239.9 F 12.5  
22:00 239.1 F 10.7  
24:00 235.0 F 8.2  
26:00 249.3 F 14.9  
28:00 260.5 F 21.2  
30:00 270.7 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 272.2 F 28.7  
2:00 272.6 F 29.2  
3:00 271.5 F 28.3  
4:00 272.5 F 28.8

Min 271.4 F 28.2  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.1 F 0.0  
5:00 203.5 F 0.0  
10:00 216.7 F 0.0  
15:00 228.7 F 0.0  
20:00 227.3 F 0.0  
25:00 230.4 F 0.0  
30:00 235.0 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 30:44  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:08:40  
CYCLE COMPLETE

04/04/2023 09:23 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_